

Declaración de Conformidad *Conformity Declaration*


Fabricante: TECFLUID S.A.
Manufacturer: Narcís Monturiol, 33
E 08960 Sant Just Desvern

Tipo de Equipo: Detector de nivel con automatismos
Type of Instrument: *Level detector with limit switches*

Modelo:
Series: LC40

Tipo de certificación:


Los detectores de nivel mencionados contienen en su interior un detector inductivo de la marca Pepperl & Fuchs GmbH, modelo SJ3,5N, con certificado PTB 99 ATEX 2219 X, y cuyo marcado EEx es:

 II2G EEx ia IIC T6

Por ser aparatos de seguridad intrínseca en un envolvente IP65, puede ser instalado en zona clasificada según el marcado del detector.

Type of Certification:

The mentioned level detectors contain inside one inductive detector from Pepperl & Fuchs GmbH, model SJ3,5N, with certificate PTB 99 ATEX 2219 X, and whose EEx marking is:

 II2G EEx ia IIC T6

Given that they are intrinsic safety apparatus into an enclosure with an ingress protection of IP65, they can be installed in a classified zone in accordance to the marking of the detector.

En Sant Just Desvern
Place:

Fecha: 25 de Enero de 2008
Date: *January 25, 2008*

Cargo:
Position:

(firma) (Signature)

Jordi Picazo i Navalón
(Nombre) (Full Name)

Gerente
Manager Director

